

BEST AVAILABLE COPY**PATENT ABSTRACTS OF JAPAN**

(11)Publication number : 09-033411

(43)Date of publication of application : 07.02.1997

(51)Int.Cl.

G01N 1/28
G01N 15/02
G02B 21/34

(21)Application number : 07-201282

(71)Applicant : NITTO SHOJI KK

(22)Date of filing : 14.07.1995

(72)Inventor : UEDA KEIZO

(54) PLATE FOR BLOOD INSPECTION AND MANUFACTURE THEREOF

(57)Abstract:

PROBLEM TO BE SOLVED: To provide a plastic plate which enables accurate blood inspection by making blood spread on it properly.

SOLUTION: A thin layer made of amorphous quartz is formed on the surface of a transparent plastic plate. The resin usable preferably is mainly composed of polyethylene terephthalate. The amorphous quartz layer can be formed by applying, drying and heating an organic solvent solution of polyperhydrosilazane.

LEGAL STATUS

[Date of request for examination]

[Date of sending the examiner's decision of rejection]

[Kind of final disposal of application other than the examiner's decision of rejection or application converted registration]

[Date of final disposal for application]

[Patent number]

[Date of registration]

[Number of appeal against examiner's decision of rejection]

[Date of requesting appeal against examiner's
decision of rejection]

[Date of extinction of right]

Copyright (C); 1998,2003 Japan Patent Office